



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No. CJS: 2021000371

Location: Weighbridge

Manufacturer: DINI Argeo 3590ET

Serial/Plant No: 83626665

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 17 November 2021

Next due: November 2022

Calibration Certificate of Accuracy


Business Name: A M A (Storage And Distribution) Ltd

Address: Rawmarsh Road
 Rotherham
 South Yorkshire
 S60 1RZ

Contact Name: Alan Oldershaw

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	12900				
2	12900				
3	12900				
4	12900				



As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23900	23900	0
33900	33900	0
43940	43940	0

Repeatability

43940	43940	0
-------	-------	---

Notes: Scales found to be operating within permitted tolerances.

Engineer Name:

C.Scott

C. Scott

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23900	23900	0
33900	33900	0
43940	43940	0

Repeatability

43940	43940	0
-------	-------	---

Traceable Weights Used: TR113311

Trading Standards Maximum Tolerance:	60	+/-	60	kg
--------------------------------------	----	-----	----	----

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020



West End Grange
 Epworth Water Gardens
 West End Road
 Epworth
 Doncaster
 DN9 1LA

Order no:

0293

QMF18 Issue 2 28.09.2013

Tel: 01427 875 867
 Web: www.allweighs.co.uk

Service & Repair Record Form

Service Name: A M A Storage & Distribution Ltd

Address: Rawmarsh Road
 Rotherham
 South Yorkshire

Postcode: S60 1RZ

Contact Name: Alan Oldershaw

Invoice Name: A M A Storage & Distribution Ltd

Address: Rawmarsh Road
 Rotherham
 South Yorkshire

Postcode: S601RZ

Contact Name: Alan Oldershaw

Schedule of Equipment

Maker	Model	Serial No	Other Information
DINI Argeo	3590ET	83626665	Calibration of weighbridge.
			Signed by engineer on behalf of customer due to covid.

The above equipment has been serviced/repared in accordance with the manufacturers' recommendations. By signing this form the customer authorises Allweighs Weighing Services to invoice any charges or contract premium that is now due for payment.


Date: 17 November 2021 Arr time: 13:45 Total working time on site: 30M

Total travel time: 2H Dep time: 14:15 Total charge: + VAT

Engineer Name: C. Scott On behalf of customer: A. Oldershaw

C. Scott

A. O



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No CJS:

202010156

Location:

Weighbridge

Manufacturer:

DINI Argeo 3590

Serial/Plant No:

83626665

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

24 November 2020

Next due:

November 2021

Calibration Certificate of Accuracy

Business Name: AMA (Storage and Distribution) Ltd

Address: Rawmarsh Road
 Rotherham
 South Yorkshire
 S60 1RZ

Contact Name: Alan Oldershaw

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	8640	0			
2	8640	0			
3	8640	0			
4	8640	0			

Four loadcell

Six loadcell

Single loadcell

Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20640	20640	0
30640	30640	0
43480	43480	0

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	8000	0
20640	20640	0
30640	30640	0
43480	43480	0

Repeatability

43480	43480	0
-------	-------	---

Repeatability

43480	43480	0
-------	-------	---

Notes:

Engineer Name:

C.Scott

C.Scott

Traceable Weights Used:

SH10932

Trading Standards Maximum Tolerance:

60 +/-

60 kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020