



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No. JS:

2023605

Location:

Weighbridge

Manufacturer:

Dini Argeo 3590

Serial/Plant No:

83626665

Capacity:

60000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

14 November 2023

Next due:

November 2024

Calibration Certificate of Accuracy

Business Name:

AMA Storage & Distribution Ltd

Address:

Grafton Bridge
 Rotherham
 S60 1RZ

Contact Name:

John Cox

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	10000				
2	10000				
3	10000				
4	10000				

Four loadcell Six loadcell

Single loadcell Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23360	23360	0
33360	33360	0
44020	44000	20

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
23360	23360	0
33360	33360	0
44020	44020	0

Repeatability

44020	44000	20
-------	-------	----

Repeatability

44020	44020	0
-------	-------	---

Notes:

Weighbridge tested and found to be accurate.

Engineer Name:

J Scott

Traceable Weights Used:

W23017 + W23018

Trading Standards Maximum Tolerance:

60

+/-

60

kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020